

Hellenic Accreditation System



ACCREDITATION CERTIFICATE

No. 421-3

The Hellenic Accreditation System (ESYD), as the national accreditation body of Greece in accordance with the Laws 3066/2002 and 4109/2013,


ACCREDITS

the
Calibration Laboratory
of
HELLAMCO S.A.
in Chalandri, Attiki, Greece

under the terms of the ELOT EN ISO/IEC 17025: 2005 Standard and the ESYD Criteria, to carry out measurements, as specified in the attached Scope of the Accreditation, which may be revised by decisions of ESYD.

The initial accreditation was issued on April 17th, 2008. This Certificate renews the accreditation and it is valid until April 17th, 2020, provided that the accredited body will comply with the above Standard and the ESYD Criteria.

Athens, September 16th, 2016



I. Sitaras

Laboratories Accreditation Division Manager

Hellenic Accreditation System



Annex F2/12 to the Certificate No. 421-3

SCOPE of ACCREDITATION

of the
Calibration Laboratory
of
HELLAMCO A.E.

Measurand / Calibration item	Range of measurement	Calibration & Measurement Capability (k=2)*	Remarks	
Mass Measurements				
Mass / Non-automatic electronic weighing instruments	1 mg	4,0 µg	Calibration is performed according to the EURAMET/cg-18/v.04-2015	
	2 mg	4,0 µg		
	5 mg	4,0 µg		
	10 mg	5,0 µg		
	20 mg	7,0 µg		Use of Weight Standards according to OIML :
	50 mg	8,0 µg		
	100 mg	11 µg		
	200 mg	13 µg		
	500 mg	17 µg		
	1 g	20 µg		
	2 g	26 µg		
	5 g	33 µg		
	10 g	40 µg		
	20 g	53 µg		
	50 g	66 µg		
	100 g	0,11 mg	Κλάσης F1 in the range: 1 g - 5 kg Maximum total available weight: 13,2 kg.	
	200 g	0,20 mg		
	500 g	0,53 mg		
	1 kg	1,06 mg		
	2 kg	6,61 mg		
	5 kg	16,5 mg		
10 kg	0,33 g	Κλάσης M1 in the range: 10 kg - 20 kg Maximum total available weight: 63,2 kg		
20 kg	0,66 g			
50 kg	1,65 g			
60 kg	1,69 g			
			Calibration is performed on-site.	

** Where uncertainty is accompanied by the corresponding unit, it is absolute, while where it is not accompanied by a unit, it is relative..*

Site of assessment: **Permanent laboratory premises, Marathonos Str. 7, Chalandri, 196 00 Attiki, Greece.**

Approved Signatories: **I. Drakos, N. Denis, D. Valilis.**

This Scope of Accreditation replaces the previous one dated 16.09. 2016.

The Accreditation Certificate No. **421-3**, to EAOT EN ISO/IEC 17025: 2005, is valid until 16.04.2020.

Athens, April 27th, 2018



Konstantinos Voutsinas
Managing Director of ESYD