



Annex F2/15 to the Certificate No. 421-3

SCOPE of ACCREDITATION

of the
Calibration Laboratory
of
HELLAMCO A.E.

Measurand / Calibration item	Range of measurement	Calibration & Measurement Capability (k=2)*	Remarks
Mass Measurements			
Mass / Non-automatic electronic weighing instruments	1 mg	4,0 µg	Calibration is performed according to the EURAMET/cg-18/v.04-2015 Use of Weight Standards according to OIML : Class E2 in the range: 1 mg - 1 kg Maximum total available weight: 2,1 kg. Κλάσης F1 in the range: 1 g - 5 kg Maximum total available weight: 13,2 kg. Κλάσης M1 in the range: 10 kg - 20 kg Maximum total available weight: 63,2 kg
	2 mg	4,0 µg	
	5 mg	4,0 µg	
	10 mg	5,0 µg	
	20 mg	7,0 µg	
	50 mg	8,0 µg	
	100 mg	11 µg	
	200 mg	13 µg	
	500 mg	17 µg	
	1 g	20 µg	
	2 g	26 µg	
	5 g	33 µg	
	10 g	40 µg	
	20 g	53 µg	
	50 g	66 µg	
	100 g	0,11 mg	
	200 g	0,20 mg	
	500 g	0,53 mg	
	1 kg	1,06 mg	
	2 kg	6,61 mg	
	5 kg	16,5 mg	
10 kg	0,33 g	Calibration is performed on-site.	
20 kg	0,66 g		
50 kg	1,65 g		
60 kg	1,69 g		

** Where uncertainty is accompanied by the corresponding unit, it is absolute, while where it is not accompanied by a unit, it is relative..*

Site of assessment: **Permanent laboratory premises, Delfon Str. 3, Chalandri, 196 00 Attiki, Greece.**

Approved Signatories: **N. Denis, D. Valilis, P. Kostrivas.**

This Scope of Accreditation replaces the previous one dated 28.01.2020

The Accreditation Certificate No. **421-3**, to ΕΛΟΤ EN ISO/IEC 17025: 2005, has been expanded by ESYD's resolution until 16.10.2020.

Athens, 30.03.2020



Spyridon Podaras
CEO of ESYD